
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 003120
16:15 08/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:01
PRESSURIZING 3:02
STERILIZING 4:03
136.4 C 220kPa 4:03
136.3 C 221kPa 4:33
136.4 C 222kPa 5:03
136.5 C 226kPa 5:33
136.2 C 224kPa 6:03
136.2 C 224kPa 6:33
136.4 C 223kPa 7:03
136.3 C 223kPa 7:33
136.3 C 223kPa 7:33
VENTING 7:33
AIR DRYING 8:32
CYCLE COMPLETE 8:38

Digital Signature #
F6D4E524D145E0E0
