

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3019

Name: RITTER 1  
Sterilizer ID: V2178198  
03/20/2023 11:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 195.5 F 0.0  
2:00 186.8 F 3.9  
4:00 192.0 F 8.2  
6:00 212.5 F 12.9  
8:00 228.0 F 8.5  
10:00 240.9 F 11.5  
12:00 239.9 F 11.4  
14:00 238.6 F 9.8  
16:00 252.6 F 16.7  
18:00 264.5 F 23.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.4  
1:00 272.4 F 29.2  
2:00 272.1 F 28.6  
3:00 271.6 F 28.7  
4:00 272.5 F 29.0  
  
Min 271.3 F 28.3  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.6 F 0.0  
5:00 224.6 F 0.0  
10:00 229.2 F 0.0  
15:00 234.6 F 0.0  
20:00 232.6 F 0.0  
25:00 221.1 F 0.0  
30:00 218.1 F 0.0  
Drying Complete

FILLING: 1:19  
HEATING: 19:46  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:19  
CYCLE COMPLETE

03/20/2023 12:43 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_