
Midmark M11 - v1.0.5
Total Cycles: 999

Name: RITTER2
Sterilizer ID: V2313228
04/18/2022 01:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.1 F 0.0
2:00 200.8 F 1.6
4:00 202.3 F 7.3
6:00 220.3 F 13.1
8:00 231.1 F 9.5
10:00 236.5 F 9.2
12:00 242.6 F 11.5
14:00 247.1 F 14.2
16:00 260.1 F 21.4
18:00 271.6 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.1 F 29.1
2:00 272.5 F 29.6
3:00 271.6 F 29.1
4:00 272.2 F 29.2

Min 271.5 F 28.8
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.3 F 0.0
5:00 211.8 F 0.0
10:00 224.0 F 0.0
15:00 226.0 F 0.0
20:00 221.4 F 0.0
25:00 218.2 F 0.0
30:00 216.2 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 18:32
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:20
CYCLE COMPLETE

04/18/2022 02:39 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____