

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000998  
13:32 16/12/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:10  
PRESSURIZING 2:10  
STERILIZING 3:13  
136.4 C 223kPa 3:13  
136.3 C 225kPa 3:43  
136.4 C 217kPa 4:13  
136.4 C 221kPa 4:43  
136.4 C 222kPa 5:13  
136.4 C 222kPa 5:43  
136.4 C 220kPa 6:13  
136.4 C 222kPa 6:43  
136.4 C 222kPa 6:43  
VENTING 6:43  
AIR DRYING 7:33  
CYCLE COMPLETE 7:34

-----  
Digital Signature #  
AF7A294B2A5FFE09  
-----