
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000611
12:25 27/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:48
STERILIZING 3:51
136.0 C 227kPa 3:51
136.2 C 231kPa 4:21
136.2 C 231kPa 4:51
136.2 C 229kPa 5:21
136.4 C 231kPa 5:51
136.4 C 233kPa 6:21
136.3 C 232kPa 6:51
136.3 C 232kPa 7:21
136.3 C 232kPa 7:21
VENTING 7:21
AIR DRYING 8:13
CASSETTE REMOVED
CYCLE COMPLETE 8:15

Digital Signature #
4459AA30210D8526
