
Midmark M11 - v1.0.0
Total Cycles: 1189

Name: RITTER 1
Sterilizer ID: V2178198
09/21/2021 08:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 125.5 F 0.0
2:00 124.9 F 0.0
4:00 137.1 F 0.2
6:00 166.1 F 1.8
8:00 205.0 F 5.5
10:00 215.1 F 9.5
12:00 225.3 F 11.6
14:00 227.2 F 7.3
16:00 236.6 F 11.8
18:00 235.5 F 8.4
20:00 244.6 F 12.3
22:00 244.4 F 12.2
24:00 255.8 F 18.3
26:00 266.3 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.4 F 29.0
2:00 272.0 F 28.5
3:00 271.3 F 28.2
4:00 272.4 F 29.0

Min 271.3 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 197.0 F 0.0
10:00 193.7 F 0.0
15:00 196.0 F 0.0
20:00 195.6 F 0.0
25:00 200.5 F 0.0
30:00 205.2 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 27:45
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:05:53
CYCLE COMPLETE

09/21/2021 09:36 AM
ACCEPTED: X
REJECTED:
Operator: U01