

-----  
Midmark M11 - v1.0.5  
Total Cycles: 158

Name: RITTER2  
Sterilizer ID: V2313228  
07/05/2021 09:14 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 147.8 F 0.0  
2:00 148.5 F 2.1  
4:00 166.5 F 5.9  
6:00 190.7 F 10.7  
8:00 209.5 F 4.2  
10:00 226.1 F 9.6  
12:00 233.1 F 9.2  
14:00 242.1 F 12.9  
16:00 242.5 F 12.0  
18:00 255.2 F 18.6  
20:00 266.5 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.3 F 29.4  
2:00 271.2 F 29.0  
3:00 272.3 F 29.5  
4:00 271.3 F 28.8

Min 271.2 F 28.7  
Max 272.5 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.8 F 0.0  
5:00 215.9 F 0.0  
10:00 221.2 F 0.0  
15:00 218.0 F 0.0  
20:00 215.1 F 0.0  
25:00 209.8 F 0.0  
30:00 204.0 F 0.0  
Drying Complete

FILLING: 1:14  
HEATING: 21:33  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:59:55  
CYCLE COMPLETE

07/05/2021 10:14 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01