

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1174

Name: RITTER 1  
Sterilizer ID: V2178198  
09/15/2021 08:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 144.7 F 0.0  
2:00 142.2 F 0.7  
4:00 147.6 F 2.4  
6:00 193.1 F 5.3  
8:00 198.9 F 9.9  
10:00 219.4 F 7.7  
12:00 223.2 F 7.9  
14:00 235.3 F 12.5  
16:00 236.6 F 9.5  
18:00 237.6 F 9.6  
20:00 247.1 F 13.7  
22:00 258.9 F 20.1  
24:00 269.5 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.5 F 29.3  
2:00 271.6 F 28.9  
3:00 272.6 F 29.3  
4:00 271.6 F 28.7

Min 271.5 F 28.6  
Max 273.3 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.5 F 0.0  
5:00 232.0 F 0.0  
10:00 219.6 F 0.0  
15:00 204.0 F 0.0  
20:00 190.6 F 0.0  
25:00 195.3 F 0.0  
30:00 201.6 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 25:54  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:04:15  
CYCLE COMPLETE

09/15/2021 09:14 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01