
Midmark M11 - v1.0.5
Total Cycles: 2196

Name: RITTER2
Sterilizer ID: V2313228
03/30/2023 08:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.3 F 0.0
2:00 158.6 F 2.2
4:00 165.7 F 5.4
6:00 185.3 F 9.9
8:00 210.2 F 5.5
10:00 221.2 F 8.9
12:00 233.8 F 10.3
14:00 239.0 F 11.7
16:00 239.0 F 10.4
18:00 252.4 F 17.2
20:00 264.0 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.6 F 29.7
2:00 271.3 F 29.1
3:00 272.4 F 29.8
4:00 271.6 F 29.2

Min 271.2 F 29.0
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.9 F 0.0
5:00 211.7 F 0.0
10:00 221.4 F 0.0
15:00 226.2 F 0.0
20:00 230.0 F 0.0
25:00 229.8 F 0.0
30:00 222.6 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 21:56
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:13
CYCLE COMPLETE

03/30/2023 09:48 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____