
Midmark M11 - v1.0.5
Total Cycles: 3237

Name: RITTER2
Sterilizer ID: V2313228
12/20/2023 12:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.0 F 0.0
2:00 175.9 F 3.2
4:00 199.5 F 8.6
6:00 219.5 F 8.8
8:00 232.7 F 10.0
10:00 232.9 F 8.0
12:00 242.2 F 11.4
14:00 246.9 F 13.8
16:00 259.9 F 21.0
18:00 271.4 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.3 F 29.0
2:00 272.4 F 29.5
3:00 271.8 F 28.6
4:00 272.6 F 29.4

Min 271.4 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.7 F 0.0
5:00 223.0 F 0.0
10:00 201.6 F 0.0
15:00 199.9 F 0.0
20:00 198.7 F 0.0
25:00 210.6 F 0.0
30:00 217.4 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 18:36
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:31
CYCLE COMPLETE

12/20/2023 01:21 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____