
Midmark M11 - v1.0.0
Total Cycles: 760

Name: RITTER 1
Sterilizer ID: V2178198
03/31/2021 10:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.6 F 0.0
2:00 180.7 F 2.1
4:00 177.2 F 6.4
6:00 202.6 F 11.7
8:00 216.2 F 5.7
10:00 231.0 F 11.3
12:00 234.8 F 8.4
14:00 240.0 F 10.1
16:00 246.6 F 13.5
18:00 259.3 F 20.4
20:00 270.6 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 271.8 F 28.6
2:00 271.7 F 28.6
3:00 272.0 F 28.5
4:00 272.2 F 28.6

Min 271.6 F 28.5
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.0 F 0.0
5:00 240.5 F 0.0
10:00 230.8 F 0.0
15:00 221.6 F 0.0
20:00 220.0 F 0.0
25:00 223.0 F 0.0
30:00 222.4 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 21:17
STERILIZING: 4:00
VENTING: 2:57
DRYING: 30:00
TOTAL CYCLE: 00:59:21
CYCLE COMPLETE

03/31/2021 11:53 AM
ACCEPTED: X
REJECTED:
Operator: U01