
Midmark M11 - v1.0.5
Total Cycles: 535

Name: RITTER2
Sterilizer ID: V2313228
11/10/2021 12:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.1 F 0.0
2:00 197.3 F 3.8
4:00 201.0 F 9.1
6:00 218.6 F 8.4
8:00 233.6 F 10.5
10:00 234.1 F 7.9
12:00 241.9 F 11.5
14:00 245.8 F 13.5
16:00 258.0 F 20.2
18:00 268.7 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.6 F 29.6
2:00 271.5 F 29.1
3:00 272.7 F 29.6
4:00 271.5 F 28.9

Min 271.4 F 28.8
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.0 F 0.0
5:00 230.8 F 0.0
10:00 229.4 F 0.0
15:00 228.9 F 0.0
20:00 228.3 F 0.0
25:00 228.7 F 0.0
30:00 229.2 F 0.0

Drying Complete

FILLING: 2:14
HEATING: 19:08
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:31
CYCLE COMPLETE

11/10/2021 01:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____