
Midmark M11 - v1.0.0
Total Cycles: 3802

Name: RITTER 1
Sterilizer ID: V2178198
09/20/2023 02:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.7 F 0.0
2:00 166.3 F 2.3
4:00 189.4 F 5.9
6:00 204.4 F 10.8
8:00 218.8 F 5.4
10:00 227.2 F 10.2
12:00 232.6 F 8.0
14:00 242.1 F 12.5
16:00 241.4 F 11.1
18:00 253.9 F 17.4
20:00 264.6 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.5 F 28.7
2:00 272.1 F 29.1
3:00 271.5 F 28.4
4:00 272.4 F 29.0

Min 271.2 F 28.3
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.3 F 0.0
5:00 221.1 F 0.0
10:00 236.2 F 0.0
15:00 233.7 F 0.0
20:00 226.5 F 0.0
25:00 227.2 F 0.0
30:00 228.5 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 21:54
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:00:51
CYCLE COMPLETE

09/20/2023 03:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____