

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 007698  
7:13 01/08/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:55  
PRESSURIZING 2:56  
STERILIZING 3:53  
136.4 C 221kPa 3:53  
136.1 C 221kPa 4:23  
136.4 C 223kPa 4:53  
136.4 C 218kPa 5:23  
136.4 C 221kPa 5:53  
136.3 C 222kPa 6:23  
135.9 C 217kPa 6:53  
136.3 C 222kPa 7:23  
136.3 C 222kPa 7:23  
VENTING 7:23  
AIR DRYING 8:13  
CYCLE COMPLETE 38:13

-----  
Digital Signature #  
978AAD7C70B02212  
-----