

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2984

Name: RITTER 1  
Sterilizer ID: V2178198  
03/08/2023 05:34 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 164.2 F 0.0  
2:00 167.2 F 3.3  
4:00 188.7 F 7.4  
6:00 208.3 F 12.5  
8:00 220.5 F 6.2  
10:00 233.0 F 11.1  
12:00 232.6 F 7.8  
14:00 243.4 F 11.8  
16:00 245.7 F 12.9  
18:00 258.5 F 19.8  
20:00 269.8 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.9 F 28.9  
2:00 271.6 F 28.0  
3:00 272.7 F 29.0  
4:00 272.2 F 28.5  
  
Min 271.5 F 28.0  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.5 F 0.0  
5:00 226.5 F 0.0  
10:00 229.3 F 0.0  
15:00 229.4 F 0.0  
20:00 228.3 F 0.0  
25:00 228.7 F 0.0  
30:00 227.2 F 0.0  
Drying Complete

FILLING: 1:10  
HEATING: 20:57  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 00:59:24  
CYCLE COMPLETE

03/08/2023 06:33 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_