
Midmark M11 - v1.0.0
Total Cycles: 2265

Name: RITTER 1
Sterilizer ID: V2178198
08/30/2022 10:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.1 F 0.0
2:00 174.1 F 1.1
4:00 189.9 F 5.1
6:00 211.5 F 10.5
8:00 220.6 F 5.5
10:00 235.2 F 11.0
12:00 233.6 F 8.0
14:00 242.5 F 11.1
16:00 245.9 F 13.2
18:00 258.6 F 20.1
20:00 269.9 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.4 F 29.0
2:00 271.5 F 28.6
3:00 272.5 F 29.0
4:00 271.4 F 28.4

Min 271.4 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.2 F 0.0
5:00 200.9 F 0.0
10:00 218.1 F 0.0
15:00 219.1 F 0.0
20:00 220.6 F 0.0
25:00 222.4 F 0.0
30:00 223.8 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 20:55
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:38
CYCLE COMPLETE

08/30/2022 11:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____