
Midmark M11 - v1.0.5
Total Cycles: 2962

Name: RITTER2
Sterilizer ID: V2313228
10/03/2023 09:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.0 F 0.0
2:00 72.3 F 0.0
4:00 97.2 F 0.0
6:00 135.0 F 0.0
8:00 172.3 F 0.5
10:00 191.3 F 1.5
12:00 204.4 F 5.7
14:00 217.4 F 9.7
16:00 228.7 F 10.4
18:00 233.0 F 9.2
20:00 240.6 F 11.4
22:00 242.7 F 12.0
24:00 241.0 F 10.8
26:00 254.6 F 17.7
28:00 265.9 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.5
1:00 271.9 F 28.8
2:00 272.4 F 29.2
3:00 272.7 F 29.3
4:00 272.8 F 29.2

Min 271.7 F 28.4
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 231.0 F 0.0
Dry Cycle Aborted

FILLING: 1:03
HEATING: 29:29
STERILIZING: 4:00
VENTING: 2:49
DRYING: 1:19
TOTAL CYCLE: 00:38:40

10/03/2023 10:17 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____