

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2987

Name: RITTER2  
Sterilizer ID: V2313228  
10/10/2023 10:53 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 163.4 F 0.0  
2:00 166.5 F 2.7  
4:00 189.5 F 7.0  
6:00 207.5 F 12.4  
8:00 221.3 F 7.5  
10:00 235.2 F 12.8  
12:00 237.7 F 10.7  
14:00 237.2 F 9.7  
16:00 251.9 F 16.9  
18:00 264.3 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.2  
1:00 272.7 F 29.7  
2:00 271.6 F 29.4  
3:00 272.5 F 29.6  
4:00 271.5 F 29.2  
  
Min 271.5 F 29.0  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.4 F 0.0  
5:00 219.4 F 0.0  
10:00 207.6 F 0.0  
15:00 204.0 F 0.0  
20:00 202.2 F 0.0  
25:00 199.2 F 0.0  
30:00 212.4 F 0.0  
Drying Complete

FILLING: 1:46  
HEATING: 19:52  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:49  
CYCLE COMPLETE

10/10/2023 11:51 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_