

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4565

Name: RITTER 1  
Sterilizer ID: V2178198  
05/17/2024 10:01 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.4 F 0.0  
2:00 71.8 F 0.0  
4:00 104.9 F 0.2  
6:00 124.3 F 2.5  
8:00 151.9 F 5.6  
10:00 173.3 F 9.6  
12:00 190.9 F 9.5  
14:00 209.5 F 4.4  
16:00 221.2 F 8.0  
18:00 232.4 F 12.1  
20:00 231.7 F 7.7  
22:00 244.4 F 13.2  
24:00 242.0 F 11.1  
26:00 255.1 F 17.7  
28:00 266.4 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.0  
1:00 272.4 F 28.3  
2:00 272.5 F 28.8  
3:00 271.5 F 27.9  
4:00 272.7 F 28.5

Min 271.4 F 27.8  
Max 272.8 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.4 F 0.0  
5:00 190.0 F 0.0  
10:00 201.4 F 0.0  
15:00 208.2 F 0.0  
20:00 210.7 F 0.0  
25:00 210.9 F 0.0  
30:00 211.6 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 29:33  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 01:07:27  
CYCLE COMPLETE

05/17/2024 11:08 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_