
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001125
12:03 18/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:25
PRESSURIZING 3:26
STERILIZING 4:34
136.4 C 220kPa 4:34
136.3 C 223kPa 5:04
136.2 C 223kPa 5:34
136.4 C 222kPa 6:04
136.4 C 222kPa 6:34
136.1 C 221kPa 7:04
136.4 C 224kPa 7:34
136.4 C 223kPa 8:04
136.4 C 223kPa 8:04
VENTING 8:04
AIR DRYING 9:05
CYCLE COMPLETE 9:14

Digital Signature #
8D9F2A70E2B68953
