

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
1.2uS / 0.7ppm  
CYCLE NUMBER 000595  
14:13 20/08/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:44  
PRESSURIZING 2:53  
STERILIZING 4:01  
136.5 C 221kPa 4:01  
136.3 C 231kPa 4:31  
136.4 C 232kPa 5:01  
136.2 C 231kPa 5:31  
136.3 C 232kPa 6:01  
136.3 C 231kPa 6:31  
136.3 C 231kPa 7:01  
136.2 C 230kPa 7:31  
136.2 C 230kPa 7:31  
VENTING 7:31  
AIR DRYING 8:22  
CASSETTE REMOVED  
CYCLE COMPLETE 8:35  
-----

Digital Signature #  
79C83CBA4B0E4490  
-----