

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4083

Name: RITTER2  
Sterilizer ID: V2313228  
11/04/2024 05:34 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 172.8 F 0.0  
2:00 174.3 F 3.5  
4:00 198.6 F 8.6  
6:00 216.0 F 10.7  
8:00 229.9 F 8.6  
10:00 238.4 F 11.1  
12:00 242.0 F 12.2  
14:00 241.7 F 11.4  
16:00 255.4 F 18.4  
18:00 267.3 F 25.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.7  
1:00 272.6 F 29.0  
2:00 272.4 F 29.5  
3:00 272.3 F 28.8  
4:00 272.7 F 29.5  
  
Min 271.7 F 28.6  
Max 273.0 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.1 F 0.0  
5:00 245.3 F 0.0  
10:00 232.0 F 0.0  
15:00 224.5 F 0.0  
20:00 214.4 F 0.0  
25:00 206.5 F 0.0  
30:00 198.9 F 0.0  
Drying Complete

FILLING: 2:00  
HEATING: 19:20  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 00:58:45  
CYCLE COMPLETE

11/04/2024 06:32 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_