

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1466

Name: RITTER2  
Sterilizer ID: V2313228  
09/08/2022 06:48 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.2 F 0.0  
2:00 80.6 F 0.0  
4:00 89.2 F 0.0  
6:00 176.5 F 0.0  
8:00 176.9 F 0.0  
10:00 189.0 F 2.4  
12:00 204.7 F 5.9  
14:00 225.6 F 9.5  
16:00 234.4 F 13.1  
18:00 229.1 F 7.7  
20:00 238.3 F 12.0  
22:00 234.2 F 9.0  
24:00 241.3 F 10.5  
26:00 245.0 F 13.2  
28:00 256.7 F 19.4  
30:00 267.5 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.0  
1:00 271.9 F 29.6  
2:00 272.8 F 29.7  
3:00 272.7 F 29.8  
4:00 272.8 F 29.9

Min 271.0 F 28.9  
Max 273.1 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.7 F 0.0  
5:00 231.7 F 0.0  
10:00 215.4 F 0.0  
15:00 213.2 F 0.0  
20:00 209.0 F 0.0  
25:00 208.4 F 0.0  
30:00 204.1 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 31:19  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 01:09:22  
CYCLE COMPLETE

09/08/2022 07:57 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_