
Midmark M11 - v1.0.0
Total Cycles: 2417

Name: RITTER 1
Sterilizer ID: V2178198
10/12/2022 09:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.2 F 0.0
2:00 159.0 F 2.7
4:00 174.0 F 5.9
6:00 189.4 F 10.3
8:00 212.5 F 4.5
10:00 222.9 F 8.0
12:00 234.7 F 12.9
14:00 235.6 F 9.6
16:00 237.4 F 8.9
18:00 246.5 F 13.9
20:00 258.6 F 20.4
22:00 269.5 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.2 F 28.9
2:00 272.6 F 29.5
3:00 271.5 F 28.7
4:00 272.4 F 29.1

Min 271.5 F 28.6
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 221.4 F 0.0
5:00 230.9 F 0.0
10:00 239.5 F 0.0
15:00 228.7 F 0.0
20:00 228.8 F 0.0
25:00 227.8 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 22:57
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:01:00
CYCLE COMPLETE

10/12/2022 10:05 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____