
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001359
15:50 23/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:22
PRESSURIZING 3:28
STERILIZING 4:44
136.0 C 220kPa 4:44
136.4 C 225kPa 5:14
136.4 C 225kPa 5:44
136.1 C 222kPa 6:14
136.4 C 225kPa 6:44
136.4 C 225kPa 7:14
136.3 C 222kPa 7:44
136.4 C 224kPa 8:14
136.4 C 224kPa 8:14
VENTING 8:14
AIR DRYING 9:10
CYCLE COMPLETE 24:10

Digital Signature #
A6E1883BE3567227
