
Midmark M11 - v1.0.0
Total Cycles: 3364

Name: RITTER 1
Sterilizer ID: V2178198
06/09/2023 08:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.5 F 0.0
2:00 188.8 F 0.0
4:00 189.2 F 2.1
6:00 206.2 F 6.5
8:00 221.2 F 11.2
10:00 224.5 F 6.0
12:00 237.0 F 11.1
14:00 234.5 F 8.4
16:00 241.7 F 10.9
18:00 246.6 F 13.4
20:00 258.3 F 19.7
22:00 268.7 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.5 F 28.8
2:00 271.7 F 28.2
3:00 271.7 F 28.5
4:00 272.4 F 28.8

Min 271.3 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.4 F 0.0
5:00 203.4 F 0.0
10:00 211.1 F 0.0
15:00 210.7 F 0.0
20:00 210.4 F 0.0
25:00 211.4 F 0.0
30:00 209.6 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 23:14
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:02:03
CYCLE COMPLETE

06/09/2023 09:48 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____