
Midmark M11 - v1.0.5
Total Cycles: 805

Name: RITTER2
Sterilizer ID: V2313228
02/15/2022 03:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.0 F 0.0
2:00 177.2 F 4.0
4:00 199.0 F 9.5
6:00 219.4 F 6.0
8:00 233.2 F 11.3
10:00 235.7 F 9.4
12:00 237.9 F 9.0
14:00 251.8 F 16.4
16:00 265.1 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 273.0 F 29.3
2:00 271.9 F 28.5
3:00 272.5 F 29.4
4:00 272.9 F 29.2

Min 271.6 F 28.4
Max 273.2 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.2 F 0.0
5:00 206.4 F 0.0
10:00 211.1 F 0.0
15:00 213.5 F 0.0
20:00 214.7 F 0.0
25:00 212.7 F 0.0
30:00 211.1 F 0.0
Drying Complete

FILLING: 2:26
HEATING: 17:42
STERILIZING: 4:00
VENTING: 2:55
DRYING: 30:00
TOTAL CYCLE: 00:57:03
CYCLE COMPLETE

02/15/2022 04:56 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____