
Midmark M11 - v1.0.0
Total Cycles: 1999

Name: RITTER 1
Sterilizer ID: V2178198
06/07/2022 08:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 117.5 F 0.0
2:00 119.6 F 0.3
4:00 135.8 F 2.1
6:00 161.7 F 5.4
8:00 189.1 F 9.4
10:00 218.9 F 8.1
12:00 221.3 F 7.6
14:00 232.1 F 12.3
16:00 232.4 F 8.4
18:00 244.2 F 11.9
20:00 242.2 F 11.8
22:00 254.4 F 18.0
24:00 265.3 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.6 F 29.6
2:00 272.0 F 28.9
3:00 271.5 F 29.0
4:00 272.6 F 29.6

Min 271.2 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.8 F 0.0
5:00 194.8 F 0.0
10:00 208.6 F 0.0
15:00 211.6 F 0.0
20:00 214.0 F 0.0
25:00 217.0 F 0.0
30:00 217.8 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 25:52
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:04:04
CYCLE COMPLETE

06/07/2022 10:14 AM
ACCEPTED: X
REJECTED:
Operator: U02