
Midmark M11 - v1.0.0
Total Cycles: 1833

Name: RITTER 1
Sterilizer ID: V2178198
04/18/2022 10:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.4 F 0.0
2:00 186.9 F 3.8
4:00 214.5 F 9.2
6:00 221.8 F 6.9
8:00 234.2 F 10.2
10:00 233.9 F 7.4
12:00 245.9 F 13.1
14:00 245.1 F 12.4
16:00 258.1 F 19.4
18:00 269.8 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.5 F 28.5
2:00 271.4 F 28.3
3:00 272.5 F 28.6
4:00 271.2 F 28.1

Min 271.2 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.1 F 0.0
5:00 234.7 F 0.0
10:00 226.7 F 0.0
15:00 233.9 F 0.0
20:00 226.8 F 0.0
25:00 222.8 F 0.0
30:00 227.6 F 0.0

Drying Complete

FILLING: 1:34
HEATING: 18:55
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:39
CYCLE COMPLETE

04/18/2022 11:38 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____