

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2347

Name: RITTER 1  
Sterilizer ID: V2178198  
09/21/2022 07:40 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.4 F 0.0  
2:00 78.9 F 0.0  
4:00 84.2 F 0.0  
6:00 119.4 F 1.2  
8:00 150.6 F 3.9  
10:00 174.2 F 7.0  
12:00 190.0 F 11.1  
14:00 213.2 F 3.5  
16:00 219.4 F 7.2  
18:00 230.5 F 11.3  
20:00 230.9 F 7.3  
22:00 243.5 F 13.0  
24:00 241.9 F 11.3  
26:00 254.5 F 17.8  
28:00 265.6 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.4 F 29.2  
2:00 271.4 F 28.4  
3:00 272.4 F 29.3  
4:00 271.8 F 28.6

Min 271.4 F 28.4  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.5 F 0.0  
5:00 207.5 F 0.0  
10:00 226.8 F 0.0  
15:00 224.4 F 0.0  
20:00 224.2 F 0.0  
25:00 225.5 F 0.0  
30:00 233.9 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 29:46  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:07:43  
CYCLE COMPLETE

09/21/2022 08:47 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_