

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1969

Name: RITTER2  
Sterilizer ID: V2313228  
01/25/2023 08:15 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 78.7 F 0.0  
2:00 80.1 F 0.0  
4:00 90.2 F 0.0  
6:00 141.2 F 0.0  
8:00 194.9 F 0.2  
10:00 206.3 F 0.9  
12:00 215.2 F 2.9  
14:00 224.5 F 6.3  
16:00 232.7 F 9.9  
18:00 239.6 F 10.7  
20:00 238.7 F 10.8  
22:00 237.3 F 8.3  
24:00 242.2 F 10.9  
26:00 246.8 F 13.7  
28:00 258.6 F 20.2  
30:00 269.3 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.6 F 29.2  
2:00 271.5 F 28.4  
3:00 272.7 F 29.3  
4:00 272.0 F 28.6  
  
Min 271.4 F 28.3  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.1 F 0.0  
5:00 199.2 F 0.0  
10:00 193.6 F 0.0  
15:00 198.9 F 0.0  
20:00 215.0 F 0.0  
25:00 219.6 F 0.0  
30:00 220.8 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 31:05  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:09:00  
CYCLE COMPLETE

01/25/2023 09:23 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_