

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4646

Name: RITTER 1  
Sterilizer ID: V2178198  
06/17/2024 08:24 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.4 F 0.0  
2:00 74.2 F 0.0  
4:00 98.2 F 0.1  
6:00 121.7 F 0.5  
8:00 144.1 F 1.9  
10:00 170.5 F 5.0  
12:00 186.2 F 8.5  
14:00 201.5 F 11.7  
16:00 212.9 F 3.2  
18:00 222.1 F 7.1  
20:00 231.7 F 10.8  
22:00 232.2 F 7.6  
24:00 241.7 F 11.9  
26:00 238.0 F 9.5  
28:00 250.6 F 15.5  
30:00 261.0 F 21.3  
32:00 270.7 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.6 F 28.8  
2:00 271.5 F 28.2  
3:00 272.7 F 28.8  
4:00 271.6 F 28.1  
  
Min 271.5 F 28.0  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.5 F 0.0  
5:00 199.9 F 0.0  
10:00 192.7 F 0.0  
15:00 214.2 F 0.0  
20:00 211.8 F 0.0  
25:00 213.7 F 0.0  
30:00 216.6 F 0.0  
  
Drying Complete

FILLING: 0:46  
HEATING: 32:48  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:11:01  
CYCLE COMPLETE

06/17/2024 09:34 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_