
Midmark M11 - v1.0.0
Total Cycles: 670

Name: MIDMARK
Sterilizer ID: V2178198
02/17/2021 10:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.1 F 0.0
2:00 163.8 F 2.8
4:00 193.4 F 7.4
6:00 220.0 F 11.7
8:00 226.7 F 9.1
10:00 236.5 F 8.7
12:00 247.0 F 13.3
14:00 244.9 F 12.2
16:00 258.4 F 19.4
18:00 270.1 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 272.6 F 28.4
1:00 271.9 F 28.7
2:00 272.2 F 28.6
3:00 272.2 F 28.3
4:00 272.1 F 28.8

Min 271.6 F 28.3
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.4 F 0.0
5:00 234.7 F 0.0
10:00 224.1 F 0.0
15:00 231.3 F 0.0
20:00 183.8 F 0.0
Dry Cycle Aborted

FILLING: 1:33
HEATING: 19:15
STERILIZING: 4:00
VENTING: 2:56
DRYING: 23:10
TOTAL CYCLE: 00:50:53

02/17/2021 10:55 AM
ACCEPTED: X
REJECTED:
Operator: U01