
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002064
8:49 20/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:51
PRESSURIZING 1:58
STERILIZING 3:03
136.5 C 225kPa 3:03
136.3 C 225kPa 3:33
136.4 C 226kPa 4:03
136.4 C 223kPa 4:33
136.4 C 224kPa 5:03
136.1 C 223kPa 5:33
136.2 C 224kPa 6:03
136.1 C 223kPa 6:33
136.1 C 223kPa 6:33
VENTING 6:33
AIR DRYING 7:24
CYCLE COMPLETE 14:25

Digital Signature #
0F53ABBCEE6C0AE9
