

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 006441  
9:59 11/10/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:02  
PRESSURIZING 3:02  
STERILIZING 4:01  
136.4 C 222kPa 4:01  
136.4 C 225kPa 4:31  
136.4 C 221kPa 5:01  
136.4 C 223kPa 5:31  
136.4 C 224kPa 6:01  
136.3 C 223kPa 6:31  
136.3 C 224kPa 7:01  
136.1 C 221kPa 7:31  
136.1 C 221kPa 7:31  
VENTING 7:31  
AIR DRYING 8:25  
CYCLE COMPLETE 8:35

-----  
Digital Signature #  
8A5762027961F8BF  
-----