
Midmark M11 - v1.0.5
Total Cycles: 2629

Name: RITTER2
Sterilizer ID: V2313228
07/11/2023 07:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.6 F 0.0
2:00 71.1 F 0.0
4:00 79.1 F 0.0
6:00 111.1 F 0.0
8:00 170.2 F 0.0
10:00 201.0 F 2.9
12:00 203.8 F 6.3
14:00 213.4 F 9.8
16:00 234.2 F 13.5
18:00 228.4 F 7.7
20:00 234.5 F 11.7
22:00 232.8 F 8.1
24:00 244.3 F 13.3
26:00 241.6 F 11.6
28:00 253.2 F 17.4
30:00 263.5 F 23.5
32:00 272.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.3
1:00 272.0 F 29.6
2:00 272.2 F 29.8
3:00 272.4 F 29.8
4:00 272.4 F 29.8

Min 271.7 F 29.2
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.9 F 0.0
5:00 228.7 F 0.0
10:00 223.9 F 0.0
15:00 225.7 F 0.0
20:00 220.4 F 0.0
25:00 228.1 F 0.0
30:00 218.1 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 32:53
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:10:54
CYCLE COMPLETE

07/11/2023 08:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____