
Midmark M11 - v1.0.5
Total Cycles: 363

Name: RITTER2
Sterilizer ID: V2313228
09/14/2021 02:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 96.2 F 0.0
2:00 107.5 F 0.7
4:00 156.9 F 2.7
6:00 147.1 F 5.4
8:00 174.7 F 8.9
10:00 193.4 F 12.6
12:00 210.6 F 4.1
14:00 223.2 F 8.0
16:00 233.0 F 12.8
18:00 235.5 F 10.2
20:00 236.5 F 9.2
22:00 247.7 F 14.7
24:00 258.6 F 20.8
26:00 268.8 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.2
1:00 272.4 F 29.6
2:00 272.3 F 29.9
3:00 271.5 F 29.2
4:00 272.2 F 29.6

Min 271.5 F 29.1
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.6 F 0.0
5:00 232.7 F 0.0
10:00 225.2 F 0.0
15:00 206.5 F 0.0
20:00 174.1 F 0.0
25:00 171.8 F 0.0
30:00 167.2 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 26:59
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:05:11
CYCLE COMPLETE

09/15/2021 06:07 AM
ACCEPTED: X
REJECTED:
Operator: U01