
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002668
10:11 08/04/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:11
PRESSURIZING 3:12
STERILIZING 4:16
136.4 C 224kPa 4:16
135.9 C 222kPa 4:46
136.3 C 225kPa 5:16
136.5 C 226kPa 5:46
136.4 C 225kPa 6:16
136.4 C 223kPa 6:46
136.4 C 225kPa 7:16
136.4 C 225kPa 7:46
136.4 C 225kPa 7:46
VENTING 7:46
AIR DRYING 8:39
CYCLE COMPLETE 38:39

Digital Signature #
A72AF8A58C57BDCE
