
Midmark M11 - v1.0.0
Total Cycles: 4329

Name: RITTER 1
Sterilizer ID: V2178198
02/23/2024 03:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.2 F 0.0
2:00 186.0 F 3.5
4:00 220.9 F 8.4
6:00 220.5 F 8.7
8:00 231.5 F 8.6
10:00 240.5 F 11.1
12:00 241.9 F 11.6
14:00 239.2 F 10.1
16:00 252.7 F 16.8
18:00 264.1 F 23.4
20:00 271.1 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 272.6 F 28.8
2:00 271.1 F 28.3
3:00 272.6 F 28.9
4:00 272.0 F 28.7

Min 271.0 F 28.2
Max 273.1 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.2 F 0.0
5:00 235.5 F 0.0
10:00 221.4 F 0.0
15:00 226.1 F 0.0
20:00 224.2 F 0.0
25:00 221.3 F 0.0
30:00 216.7 F 0.0
Drying Complete

FILLING: 1:47
HEATING: 20:00
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 00:59:16
CYCLE COMPLETE

02/23/2024 04:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____