

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000791  
8:46 30/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:40  
PRESSURIZING 2:45  
STERILIZING 3:52  
136.4 C 233kPa 3:52  
136.3 C 233kPa 4:22  
136.2 C 232kPa 4:52  
136.2 C 232kPa 5:22  
136.4 C 233kPa 5:52  
136.4 C 233kPa 6:22  
136.4 C 233kPa 6:52  
136.4 C 233kPa 7:22  
136.4 C 233kPa 7:22  
VENTING 7:22  
AIR DRYING 8:14  
CYCLE COMPLETE 8:17  
-----

Digital Signature #  
B5D16078B3DA1B93  
-----