

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001571  
14:56 11/10/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:45  
PRESSURIZING 6:27  
STERILIZING 8:00  
136.0 C 220kPa 8:00  
136.2 C 221kPa 8:30  
136.2 C 221kPa 9:00  
136.4 C 220kPa 9:30  
136.4 C 220kPa 10:00  
136.3 C 222kPa 10:30  
136.1 C 220kPa 11:00  
136.2 C 221kPa 11:30  
VENTING 11:30  
AIR DRYING 12:33  
CYCLE COMPLETE 32:32

-----  
Digital Signature #  
B7B134ED98E511DC  
-----