
Midmark M11 - v1.0.0
Total Cycles: 786

Name: RITTER 1
Sterilizer ID: V2178198
04/15/2021 09:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.4 F 0.0
2:00 115.2 F 0.1
4:00 128.0 F 0.3
6:00 153.5 F 3.0
8:00 182.8 F 6.9
10:00 214.8 F 11.5
12:00 220.3 F 5.7
14:00 230.1 F 10.7
16:00 232.0 F 7.2
18:00 245.8 F 13.4
20:00 242.8 F 11.6
22:00 255.6 F 18.2
24:00 266.6 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.0 F 28.6
2:00 271.9 F 29.0
3:00 271.4 F 28.2
4:00 271.9 F 28.8

Min 271.2 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.5 F 0.0
5:00 237.1 F 0.0
10:00 231.2 F 0.0
15:00 227.3 F 0.0
20:00 220.7 F 0.0
25:00 228.1 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 25:30
STERILIZING: 4:00
VENTING: 2:57
DRYING: 30:00
TOTAL CYCLE: 01:03:23
CYCLE COMPLETE

04/15/2021 10:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____