
Midmark M11 - v1.0.0
Total Cycles: 4601

Name: RITTER 1
Sterilizer ID: V2178198
06/03/2024 09:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 106.1 F 0.0
2:00 107.1 F 0.7
4:00 113.6 F 2.4
6:00 137.2 F 4.8
8:00 165.8 F 7.6
10:00 182.4 F 11.0
12:00 210.1 F 4.8
14:00 217.0 F 6.5
16:00 226.5 F 10.1
18:00 232.9 F 9.7
20:00 236.0 F 10.2
22:00 237.5 F 9.0
24:00 245.3 F 13.1
26:00 255.9 F 18.7
28:00 265.8 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.2 F 29.0
2:00 271.9 F 29.1
3:00 271.8 F 28.7
4:00 272.3 F 29.3

Min 271.3 F 28.5
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 205.0 F 0.0
10:00 197.4 F 0.0
15:00 199.4 F 0.0
20:00 212.6 F 0.0
25:00 216.9 F 0.0
30:00 218.7 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 29:51
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:08:30
CYCLE COMPLETE

06/03/2024 11:04 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____