

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4502

Name: RITTER2  
Sterilizer ID: V2313228  
04/10/2025 11:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 150.2 F 0.0  
2:00 147.1 F 2.3  
4:00 155.7 F 5.5  
6:00 186.8 F 10.5  
8:00 210.8 F 3.8  
10:00 220.3 F 8.4  
12:00 237.6 F 13.2  
14:00 237.6 F 10.1  
16:00 236.2 F 8.6  
18:00 249.1 F 14.9  
20:00 260.7 F 21.5  
22:00 271.4 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.7  
1:00 272.9 F 30.0  
2:00 271.8 F 29.1  
3:00 273.1 F 29.9  
4:00 271.6 F 29.1

Min 271.0 F 28.6  
Max 273.4 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.8 F 0.0  
5:00 244.7 F 0.0  
10:00 225.5 F 0.0  
15:00 201.9 F 0.0  
20:00 178.2 F 0.0  
25:00 175.7 F 0.0  
30:00 172.8 F 0.0  
Drying Complete

FILLING: 1:09  
HEATING: 22:39  
STERILIZING: 4:00  
VENTING: 3:28  
DRYING: 30:00  
TOTAL CYCLE: 01:01:16  
CYCLE COMPLETE

04/10/2025 12:24 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_