
Midmark M11 - v1.0.0
Total Cycles: 751

Name: RITTER 1
Sterilizer ID: V2178198
03/25/2021 03:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.7 F 0.0
2:00 139.1 F 1.6
4:00 162.5 F 4.6
6:00 189.4 F 9.5
8:00 216.9 F 6.9
10:00 230.8 F 9.9
12:00 230.6 F 7.7
14:00 242.2 F 12.9
16:00 240.8 F 11.1
18:00 254.3 F 17.9
20:00 266.0 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.0 F 28.9
2:00 272.4 F 29.4
3:00 272.1 F 29.3
4:00 272.2 F 29.5

Min 271.0 F 28.5
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.9 F 0.0
5:00 231.8 F 0.0
10:00 235.0 F 0.0
15:00 228.5 F 0.0
20:00 230.4 F 0.0
25:00 228.5 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 21:34
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:59:51
CYCLE COMPLETE

03/29/2021 06:24 AM
ACCEPTED: X
REJECTED:
Operator: U04