

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2698

Name: RITTER2  
Sterilizer ID: V2313228  
07/25/2023 01:28 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 147.8 F 0.0  
2:00 152.1 F 0.1  
4:00 170.5 F 3.2  
6:00 189.2 F 6.9  
8:00 205.5 F 10.7  
10:00 217.9 F 6.7  
12:00 227.5 F 8.9  
14:00 237.9 F 13.4  
16:00 238.3 F 10.8  
18:00 236.6 F 9.1  
20:00 249.3 F 15.3  
22:00 260.0 F 21.3  
24:00 269.7 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.9  
1:00 272.5 F 29.6  
2:00 271.5 F 28.8  
3:00 272.4 F 29.7  
4:00 272.0 F 29.1

Min 271.3 F 28.8  
Max 272.6 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.3 F 0.0  
5:00 203.5 F 0.0  
10:00 206.3 F 0.0  
15:00 204.3 F 0.0  
20:00 200.5 F 0.0  
25:00 198.8 F 0.0  
30:00 197.1 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 25:01  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:03:44  
CYCLE COMPLETE

07/25/2023 02:31 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_