
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001631
11:55 29/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:37
STERILIZING 3:37
136.4 C 220kPa 3:37
136.2 C 224kPa 4:07
136.4 C 225kPa 4:37
136.4 C 226kPa 5:07
136.4 C 226kPa 5:37
136.4 C 225kPa 6:07
136.1 C 223kPa 6:37
136.4 C 225kPa 7:07
136.4 C 225kPa 7:07
VENTING 7:08
AIR DRYING 8:03
CYCLE COMPLETE 8:10

Digital Signature #
1AF2727A10D90081
