
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001834
9:30 22/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:20
PRESSURIZING 3:21
STERILIZING 4:29
136.0 C 220kPa 4:29
136.4 C 224kPa 4:59
136.2 C 221kPa 5:29
136.4 C 224kPa 5:59
136.4 C 223kPa 6:29
136.4 C 225kPa 6:59
136.4 C 224kPa 7:29
136.3 C 224kPa 7:59
136.3 C 224kPa 7:59
VENTING 7:59
AIR DRYING 8:55
CYCLE COMPLETE 9:13

Digital Signature #
3505BC68987341F1
