
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000485
11:36 18/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:08
PRESSURIZING 4:08
STERILIZING 5:07
136.0 C 220kPa 5:07
136.3 C 224kPa 5:37
136.4 C 225kPa 6:07
136.2 C 223kPa 6:37
136.4 C 222kPa 7:07
136.3 C 221kPa 7:37
136.4 C 223kPa 8:07
136.4 C 223kPa 8:37
136.4 C 223kPa 8:37
VENTING 8:37
AIR DRYING 9:33
CYCLE COMPLETE 12:18

Digital Signature #
C2B5172FE5293440
