

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000523  
13:08 28/05/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:06  
PRESSURIZING 3:10  
STERILIZING 4:16  
136.0 C 228kPa 4:16  
136.1 C 231kPa 4:46  
136.4 C 233kPa 5:16  
136.3 C 233kPa 5:46  
136.3 C 233kPa 6:16  
136.4 C 232kPa 6:46  
136.3 C 232kPa 7:16  
136.2 C 231kPa 7:46  
136.2 C 231kPa 7:46  
VENTING 7:46  
AIR DRYING 8:38  
CYCLE COMPLETE 22:32  
-----

Digital Signature #  
3F31BF94822F19AF  
-----